

# ***Compton Transmission Polarimeter for BNL SRF Gun***

---


5 MeV Polarimeter

Riad Suleiman

Wednesday, September 30, 2020

 Jefferson Lab

# Compton Transmission Polarimeter




---


# July 10, 2018 Test at CEBAF Injector

Dalia Lucero Ramírez Guadarrama

# Laser Table, Beamline & DAQ Schematics




# Helicity Generator Signals



# Test Setup

- Strained SupperLattice GaAs
- 780 nm, 250 MHz, Circularly polarized
- Insertable Half-wave-plate, IHWP: IN/OUT
- Beam Total Energy: 5.9 MeV
- Beam current: 0 – 3.5  $\mu\text{A}$
- Helicity Settings:
  - Clock: Free Clock
  - Delay: 0, 8 windows
  - Pattern: Quartet
  - Settle Time: 500  $\mu\text{s}$
  - Stable Time: 33330  $\mu\text{s}$
  - Frequency: 29.6 Hz



# Compton Transmission Polarimeter

- Detector:
  - Compton Transmission detector at -1.2 kV
- Magnet:



3" diameter x 2.5" long EJ-200  
Plastic Scintillator painted  
with EJ-510

3-inch  
PMT

# Asymmetries

- Charge Asymmetry:

- $I^+$  : Raw channel for +helicity
- $I^-$  : Raw channel for – helicity

$$A_{BCM} = \frac{I^+ - I^-}{I^+ + I^-}$$

- Detector Asymmetry:

- $D^+$  : Raw channel for +helicity
- $D^-$  : Raw channel for – helicity




$$A_{Det} = \frac{D^+ - D^-}{D^+ + D^-}$$

- Compton Transmission Asymmetry:

$$A_{Comp} = \frac{D^+/I^+ - D^-/I^-}{D^+/I^+ + D^-/I^-} = A_{Det} - A_{BCM}$$




# Raw Signals vs Beam Current



# Raw Signals and Asymmetries – Run 8823



- 4 Amps, OUT, 1.9  $\mu\text{A}$ , 175 sec.



Detector Rate:


$$R = \frac{1}{4 \times 0.0333 \times 0.002344^2}$$

$\sim 1.4 \text{ MHz}$




# Charge Asymmetry

- Beam current: 1.7  $\mu\text{A}$




# Detector Asymmetry vs Run Number

- Beam current:  $1.7\ \mu\text{A}$
- Magnet Current: -4 – 4 Amps




# Compton Transmission Asymmetry vs Run Number

- Beam current:  $1.7\ \mu\text{A}$
- Magnet Current: -4 – 4 Amps




# Compton Asymmetry vs Magnet Current

- Average Compton Transmission Asymmetry:  $(OUT - IN)/2$




# Compton Asymmetry vs Horizontal Wien Angle

- Beam current: 1.7  $\mu\text{A}$
- Magnet Current: +1 Amps



# Compton Asymmetry vs Beam Current

- Magnet Current: +1 Amps





---

# Design of BNL Compton Transmission Polarimeter

# GEANT4 Optimization


- Include Polarized Physics:

Particle	Process Name	Class Name (GEANT4)
Electron	Multiple Scattering	G4eMultipleScattering
	Discrete and Continuous Energy Loss	G4ePolarizedIonisation
	Bremsstrahlung	G4ePolarizedBremsstrahlung
Positron	Multiple Scattering	G4eMultipleScattering
	Discrete and Continuous Energy Loss	G4ePolarizedIonisation
	Bremsstrahlung	G4ePolarizedBremsstrahlung
	Annihilation	G4eplusPolarizedAnnihilation
Gamma	Photoelectric Effect	G4PolarizedPhotoElectricEffect
	Compton Scattering	G4PolarizedCompton
	Gamma Conversion	G4PolarizedGammaConversion

- Goal:
- Maximize  $A_p(E_b, L_{magnet}, \dots)$
- Minimize  $\Delta A$  (*i.e.*, maximize number of high energy gammas in detector)



# New Portable DAQ

- DAQ capable of :
  - Integrating two signals: PMT, BCM
  - Handling helicity signals



- Remote controllable IHWP

# Radiator





**Jefferson Lab**

Wednesday, September 30, 2020

