



PRELIMINARY
02-26-14 0600
REV 03

- NOTES:**
- ASME Y14.5-2009 IS USED FOR ALL DIMENSIONING AND TOLERANCING.
 - COIL SHALL BE FABRICATED FROM A SINGLE LENGTH OF WIRE WITHOUT JOINTS. ESTIMATED LENGTH TO BE 463 METERS OR 1518 FEET.
 - USE HYSOL RE4183 AND HD3561.
 - MAXIMUM MATERIAL CONDITION AFTER CURING.
 - CRIMP AND SOLDER LUGS ITEM 7 TO LEADS.



QTY	ITEM NO	PART OR IDENTIFYING NO	DESCRIPTION OR DISCRIPTION	MATERIAL SPECIFICATION	NOTES
2	7	MCMASTER-CARR PN 69145X215	SPADE TERMINAL VINYL INSUL 16-14 AWG #8 SCREW	-	5
1	6	-	MANUFACTURER'S NAME & DATA LABEL	-	-
2	5	-	LEAD LABEL	PRESSURE-SENSITIVE VINYL CLOTH TAPE	-
AR	4	-	TAPE ELECTRICAL FIBERGLASS NON-ADHESIVE	-	-
AR	2	-	SLEEVING SIZE NO. 13	-	-
-	1	MWS WIRE INDUSTRIES	14 AWG SQ CU WIRE HEAVY BUILD POLY IMIDE INSUL	NEMA SPEC MW 30C	2

PARTS LIST		APPROVALS		DATE	
DRAWN	S.A. OGDORY	CHECKED		06JUN14	
APPROVED		APPROVED			

UNITED STATES DEPARTMENT OF ENERGY	JEFFERSON LAB	NEGOTIATION FACILITY
MAG CORRECTOR DL MAGNET COIL		
SIZE (DWG NO)	JL0008100	REV -
SCALE	1:1	SHEET 1 OF 1